

SZABMU/PIMS Satellite lab
Fourth floor Jinnah Burn and Reconstructive Surgery Center
Allama Iqbal Medical College Lahore

Tender No. SZABMU-LP-04-2017

Dated: 13.09.2017

TENDER NOTICE

Sealed tenders / quotations are invited from manufacturers, suppliers & importers for the purchase of below listed items.

S. No.	Description	Approx. Cost	Mode of Purchase
1.	Equipments	1.8 M	FOR, SZABMU/ PIMS

Detail of Item specifications can be seen & downloaded from www.szabmu.edu.pk.

Sealed tenders addressed to the Principal Investigator should receive on **28.09.2017 13:30 Hrs.** at “SZABMU/PIMS Satellite lab Fourth floor Jinnah Burn and reconstructive surgery Center, Allama Iqbal Medical College Lahore” and will be opened by the Institutional Purchase Committee on the same day at **14:00:Hrs.** in the presence of the bidders or their authorized representatives who may wish to attend.

Administrative Officer,
Tel: 042-99232735 Ext.1525

CATEGORY EQUIPMENT			
Sr. #	Description	Detail	Qty.
1.	Refrigerated Centrifuge Machine	Centrifuge Refrigerated Max. capacity of conical tubes per rotor: 36 × 15 mL/16 × 50 mL Plate rotors for centrifugation of all types of MTP, PCR, cell culture, or Deepwell Plates Fixed angel rotors ependorf tubes High centrifugation speed of up to 20,913 × g (14,000 rpm) Aerosol-tight QuickLock Temperature range from -4 °C to 40 °C Speed 200 – 14,000 rpm Power supply 230 V, 50 – 60 Hz	1
2	Rotors for Bench top centrifuge Rotina 380R	Max. capacity of conical tubes per rotor: 36 × 15 mL/16 × 50 mL Plate rotors for centrifugation of all types of MTP, PCR, cell culture, or Deep well Plates Fixed angel rotors ependorf tubes High centrifugation speed of up to 20,913 × g (14,000 rpm)	1
2.	PCR Machine	Peltier based PCR machine with full thermal gradient which must be distributed from A to H. Sample capacity 96 x 0.2 ml tubes, 0.2 ml tube strips, or 1 x 96-well plate, Touchdown instant incubation capacity, Minimum color touch display 5.7" VGA color touch screen, Temperature range 4-100° c, gradient range 30 to 100° c. 8 temperature set points with Gradient capability, Temperature differential range 1-25° c .Temperature uniformity ± 0.5 ° c well to well within 30 sec of arrival at target temperature and accuracy ± 0.5 ° c of programmed target. Minimum programs memory 500 typical programs; unlimited with USB flash drive expansion, System must be Microsoft windows based and compatible with PC. Minimum Average ramp rate 2.5°C/sec, input 100–150 VAC, 50–60 Hz; 220–240 VAC, 50–60 Hz; 700 W maximum	2

Terms & Condition:-

1. Please quote only Ready Stock items.
2. Please mention our tender serial number on your Quotation / Tender.
3. Quotation Should Not Be Hand written.
4. Quoted rates validity should be 8 week from the date of opening.
5. Every envelope of Quotation / Tender should be marked with word “**Tender or Quotation**” and reference No.
6. Sale Tax should not be included in the quoted price.
7. Income Tax as per Government Rule.
8. Only registered suppliers who are on active Tax Payer List (ATL) of FBR eligible to supply goods/services.
9. **Earnest money at the rate of 2% of total quoted value cost in the shape of deposit at call in the name of Principal Investigator SZABMU/PIMS must accompany the tender, failing which the tender will be rejected.**

If there is any confusion about the description of material. Please come and see the undersigned.

Administrative Officer,
Tel: 042-99232735 Ext.1525